

## **UK DECLARATION OF CONFORMITY**

Manufacturer:

Mercator Medical (Thailand) Ltd.

88/8 Moo.12 Tambon Kampaengphet,

Amphur Rattaphum, Songkhla 90180 Thailand

Declare under its sole responsibility that protective gloves:

Brand	Type	Sizes	Reference Numbers
MERCATOR ideall grip+	Nitrile, powder free, black, diamond textured, for single use	S (6-7) - XXL (10-11)	RP30026002-06_0001
MERCATOR ideall grip+	Nitrile, powder free, orange, diamond textured, for single use	S (6-7) - XXL (10-11)	RP30027002-06_0001

The products described above are in conformity with the provision of the PPE Regulation (EU) 2016/425 as brought into UK law and amended as a Category III product, Type B where such is the case, with standards EN ISO 374-1:2016+A1:2018, EN 374-2:2014, EN ISO 374-4:2019, EN ISO 374-5:2016, EN 16523-1:2015+A1:2018 and EN ISO 21420:2020.

The products are identical to Personal Protective equipment, which is the subject to the UKCA Type Examination (Module B) and Module C2 On-going Conformity under certificate No. AB0321/17234-02/E00-00 issued by Approved Body:

## SATRA Technology Centre Limited (Approved Body number 0321)

Wyndham Way, Telford Way, Kettering, Northamptonshire, NN16 8SD, United Kingdom

This declaration is supported by the Quality System approval to ISO13485:2016/EN ISO13485:2016 and ISO 9001:2015 issued by:

## **SGS United Kingdom Ltd**

Rossmore Business Park Ellesmere Port Cheshire CH65 3EN UK

Date and place of issue: 12.11.2021, Thailand

Signed on the behalf of the Manufacturer:

Mr. Praneet Inthajak

Senior QA/RA Marrager (THAILAND

Mercator Medical (Thailand) Ltd.

Mercator Medical (Thailand) LTD.

88/8 Moo 12 Tambon Kampaengphet Amphur Rattaphum Songkhla 90180 Thailand ph.: +66 74 584 222 fax.: +66 74 584 223

www.mercator.co.th

fax.: +66 74 584 223 enquiry@mercator.co.th Sales Office
Mercator Medical S.A.

ul. H. Modrzejewskiej 30 31-327 Kraków, Poland ph.: +48 (12) 66 55 480 fax.: +48 (12) 66 55 415

medical@mercatormedical.eu
www.mercatormedical.eu